

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068126.D
 Acq On : 16 Aug 2021 19:28
 Operator : JC/MD
 Sample : PB138399ZHE#15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#15

Quant Time: Aug 17 07:17:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	244901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	500641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	510898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	180060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	207769	59.928	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 119.860%			
35) Dibromofluoromethane	8.024	113	152953	52.423	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.840%			
50) Toluene-d8	10.443	98	685575	50.844	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.680%			
62) 4-Bromofluorobenzene	12.733	95	221875	45.513	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 91.020%			
Target Compounds					Qvalue	
16) Acetone	4.291	43	40441	31.969	ug/l	96
18) Methyl Acetate	4.856	43	12432m	3.993	ug/l	
20) Methylene Chloride	5.095	84	28265	9.446	ug/l	90
43) Isopropyl Acetate	8.568	43	36657	5.047	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068126.D
Acq On : 16 Aug 2021 19:28
Operator : JC/MD
Sample : PB138399ZHE#15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#15

Quant Time: Aug 17 07:17:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

