

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068127.D
 Acq On : 16 Aug 2021 19:52
 Operator : JC/MD
 Sample : PB138399ZHE#16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#16

Quant Time: Aug 17 07:17:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	232065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	479646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	512991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	174851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	196377	59.775	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 119.540%			
35) Dibromofluoromethane	8.024	113	147057	52.609	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.220%			
50) Toluene-d8	10.446	98	661838	51.202	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.400%			
62) 4-Bromofluorobenzene	12.736	95	217695	46.491	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 92.980%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	39199	32.701	ug/l	93
18) Methyl Acetate	4.862	43	14750	5.026	ug/l #	78
20) Methylene Chloride	5.090	84	28890m	10.211	ug/l	
43) Isopropyl Acetate	8.563	43	34694	4.986	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068127.D
Acq On : 16 Aug 2021 19:52
Operator : JC/MD
Sample : PB138399ZHE#16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#16

Quant Time: Aug 17 07:17:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

