

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073939.D
 Acq On : 16 Aug 2022 11:07
 Operator : JC\MD
 Sample : VN0816MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0816MBL01

Quant Time: Aug 17 01:09:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	630780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	910951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	859157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	464770	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	294267	47.897	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.800%
35) Dibromofluoromethane	7.951	113	289010	53.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.300%
50) Toluene-d8	10.380	98	943862	45.238	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.480%
62) 4-Bromofluorobenzene	12.669	95	388570	56.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073939.D
Acq On : 16 Aug 2022 11:07
Operator : JC\MD
Sample : VN0816MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0816MBL01

Quant Time: Aug 17 01:09:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

