

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073940.D
 Acq On : 16 Aug 2022 11:31
 Operator : JC\MD
 Sample : VN0816WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0816WBL01

Quant Time: Aug 17 01:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	603497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	879998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	813093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	400751	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	297011	50.529	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.060%
35) Dibromofluoromethane	7.951	113	290427	55.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.620%
50) Toluene-d8	10.380	98	945569	46.914	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.668	95	368591	55.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073940.D
Acq On : 16 Aug 2022 11:31
Operator : JC\MD
Sample : VN0816WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0816WBL01

Quant Time: Aug 17 01:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

