

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073944.D
 Acq On : 16 Aug 2022 13:19
 Operator : JC\MD
 Sample : N4200-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-GW-1

Quant Time: Aug 17 01:13:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	650479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	951301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	925602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	497020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	307084	48.469	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 96.940%			
35) Dibromofluoromethane	7.951	113	300934	53.491	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.980%			
50) Toluene-d8	10.380	98	1009868	46.348	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.700%			
62) 4-Bromofluorobenzene	12.668	95	427617	59.135	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 118.260%			
Target Compounds						
16) Acetone	4.228	43	26887	8.995	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073944.D
Acq On : 16 Aug 2022 13:19
Operator : JC\MD
Sample : N4200-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-GW-1

Quant Time: Aug 17 01:13:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

