

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073954.D
 Acq On : 16 Aug 2022 17:25
 Operator : JC\MD
 Sample : PB146976ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146976ZHE#10

Quant Time: Aug 17 01:18:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	590243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	886408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	873947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	459795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	289787	50.407	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.820%	
35) Dibromofluoromethane	7.951	113	281323	53.666	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.340%	
50) Toluene-d8	10.380	98	944742	46.534	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.060%	
62) 4-Bromofluorobenzene	12.674	95	393765	58.440	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	96981	35.756	ug/l	99
20) Methylene Chloride	5.016	84	20480	3.578	ug/l	93
37) Ethyl Acetate	7.339	43	76225	8.641	ug/l	99
43) Isopropyl Acetate	8.498	43	388905	30.948	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073954.D
Acq On : 16 Aug 2022 17:25
Operator : JC\MD
Sample : PB146976ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146976ZHE#10

Quant Time: Aug 17 01:18:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

