

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073960.D
 Acq On : 16 Aug 2022 19:54
 Operator : JC\MD
 Sample : N4203-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 61023

Quant Time: Aug 17 01:21:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	726698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	1058098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	1000753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	567670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	334507	47.260	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.520%
35) Dibromofluoromethane	7.951	113	330055	52.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.500%
50) Toluene-d8	10.381	98	1100107	45.394	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.780%
62) 4-Bromofluorobenzene	12.674	95	487275	60.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	121.160%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	197638	59.185	ug/l	99
52) Toluene	10.445	92	134611	7.156	ug/l	100
67) Ethyl Benzene	11.786	91	157868	4.463	ug/l	100
68) m/p-Xylenes	11.892	106	98051	6.823	ug/l	97
69) o-Xylene	12.222	106	17667	1.226	ug/l	100
84) 1,2,4-Trimethylbenzene	13.310	105	113919	3.393	ug/l	100
95) Naphthalene	15.433	128	238715	6.004	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073960.D
Acq On : 16 Aug 2022 19:54
Operator : JC\MD
Sample : N4203-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
61023

Quant Time: Aug 17 01:21:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

