

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
 Data File : VN073962.D
 Acq On : 16 Aug 2022 20:43
 Operator : JC\MD
 Sample : N4203-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33105

Quant Time: Aug 17 01:23:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	690124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	1000472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	981897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	575137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	320952	47.748	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.500%		
35) Dibromofluoromethane	7.951	113	316492	53.491	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	106.980%		
50) Toluene-d8	10.381	98	1071860	46.776	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.560%		
62) 4-Bromofluorobenzene	12.669	95	479606	63.064	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	126.120%#		
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.293	59	8895	5.916	ug/l #	80
16) Acetone	4.222	43	463271	146.085	ug/l	97
17) Carbon Disulfide	4.457	76	24167	1.866	ug/l	96
18) Methyl Acetate	4.775	43	713983	92.232	ug/l	98
20) Methylene Chloride	5.016	84	86365	12.906	ug/l	96
25) 2-Butanone	7.263	43	183360	36.465	ug/l	100
37) Ethyl Acetate	7.340	43	93655	9.407	ug/l	99
43) Isopropyl Acetate	8.498	43	309863	21.847	ug/l #	92
52) Toluene	10.445	92	25398	1.428	ug/l	98
67) Ethyl Benzene	11.786	91	36631	1.056	ug/l	94
68) m/p-Xylenes	11.892	106	31309	2.220	ug/l	91
78) n-propylbenzene	12.863	91	88912	1.992	ug/l	99
80) 1,3,5-Trimethylbenzene	12.998	105	259716	7.624	ug/l	99
84) 1,2,4-Trimethylbenzene	13.310	105	1333067	39.188	ug/l	97
85) sec-Butylbenzene	13.439	105	158654	3.905	ug/l	78
86) p-Isopropyltoluene	13.557	119	193573	5.592	ug/l #	78
89) n-Butylbenzene	13.880	91	218927	7.985	ug/l #	76
95) Naphthalene	15.439	128	181759	4.512	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081622\
Data File : VN073962.D
Acq On : 16 Aug 2022 20:43
Operator : JC\MD
Sample : N4203-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33105

Quant Time: Aug 17 01:23:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

