

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081623\
 Data File : VN078900.D
 Acq On : 16 Aug 2023 14:04
 Operator : JC\MD
 Sample : PB154855ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154855ZHE#03

Quant Time: Aug 17 01:16:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	147084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	286612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	83870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	126481	55.262	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.520%		
35) Dibromofluoromethane	8.171	113	97534	49.633	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.260%		
50) Toluene-d8	10.571	98	376276	51.569	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.140%		
62) 4-Bromofluorobenzene	12.853	95	115429	48.399	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.800%		
Target Compounds						
						Qvalue
16) Acetone	4.447	43	28887	30.589	ug/l	98
20) Methylene Chloride	5.277	84	14301	7.106	ug/l #	77
25) 2-Butanone	7.488	43	9062	5.985	ug/l #	66
37) Ethyl Acetate	7.571	43	4406	1.332	ug/l #	77
43) Isopropyl Acetate	8.694	43	204012	34.899	ug/l #	72
51) 4-Methyl-2-Pentanone	10.453	43	40575	11.703	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081623\
Data File : VN078900.D
Acq On : 16 Aug 2023 14:04
Operator : JC\MD
Sample : PB154855ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154855ZHE#03

Quant Time: Aug 17 01:16:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

