

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081623\
 Data File : VN078902.D
 Acq On : 16 Aug 2023 14:52
 Operator : JC\MD
 Sample : PB154855ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154855ZHE#05

Quant Time: Aug 17 01:16:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	149112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	298798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	291987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	94343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	125320	54.010	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.020%	
35) Dibromofluoromethane	8.171	113	102241	50.832	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.660%	
50) Toluene-d8	10.571	98	386795	51.791	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.580%	
62) 4-Bromofluorobenzene	12.853	95	116888	47.883	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	27732	28.966	ug/l	89
20) Methylene Chloride	5.283	84	14530	7.122	ug/l #	67
25) 2-Butanone	7.494	43	9119	5.941	ug/l #	69
43) Isopropyl Acetate	8.694	43	236815	39.579	ug/l #	70
51) 4-Methyl-2-Pentanone	10.453	43	39443	11.115	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081623\
Data File : VN078902.D
Acq On : 16 Aug 2023 14:52
Operator : JC\MD
Sample : PB154855ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154855ZHE#05

Quant Time: Aug 17 01:16:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

