

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
 Data File : VN083354.D
 Acq On : 16 Aug 2024 15:15
 Operator : JC\MD
 Sample : P3456-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Aug 17 01:47:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	128651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	268108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	119438	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	107574	58.745	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.500%
35) Dibromofluoromethane	8.171	113	86809	54.616	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.240%
50) Toluene-d8	10.571	98	320673	54.086	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.180%
62) 4-Bromofluorobenzene	12.847	95	133093	57.580	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	115.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
Data File : VN083354.D
Acq On : 16 Aug 2024 15:15
Operator : JC\MD
Sample : P3456-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Aug 17 01:47:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

