

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
 Data File : VN083355.D
 Acq On : 16 Aug 2024 15:39
 Operator : JC\MD
 Sample : P3628-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 371

Quant Time: Aug 17 01:47:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	104683	59.813	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	119.620%	
35) Dibromofluoromethane	8.171	113	83099	53.715	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	107.440%	
50) Toluene-d8	10.571	98	314727	54.538	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	109.080%	
62) 4-Bromofluorobenzene	12.853	95	129417	57.524	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	115.040%	
Target Compounds						
16) Acetone	4.447	43	3370	4.921	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
Data File : VN083355.D
Acq On : 16 Aug 2024 15:39
Operator : JC\MD
Sample : P3628-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
371

Quant Time: Aug 17 01:47:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

