

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
 Data File : VN083356.D
 Acq On : 16 Aug 2024 16:03
 Operator : JC\MD
 Sample : PB162767TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162767TB

Quant Time: Aug 17 01:47:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	123984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	249922	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	276804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	106581	60.394	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 120.780%		
35) Dibromofluoromethane	8.171	113	84759	54.334	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.660%		
50) Toluene-d8	10.571	98	323305	55.560	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.120%		
62) 4-Bromofluorobenzene	12.847	95	137140	60.452	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 120.900%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	22097	31.999	ug/l	92
20) Methylene Chloride	5.283	84	7968	5.014	ug/l #	80
25) 2-Butanone	7.488	43	6942	6.547	ug/l	94
43) Isopropyl Acetate	8.694	43	148438	31.384	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
Data File : VN083356.D
Acq On : 16 Aug 2024 16:03
Operator : JC\MD
Sample : PB162767TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162767TB

Quant Time: Aug 17 01:47:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

