

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
 Data File : VN083361.D
 Acq On : 16 Aug 2024 18:05
 Operator : JC\MD
 Sample : PB162767ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162767ZHE#05

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda

Quant Time: Aug 17 01:53:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	117235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	238600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115044	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	105670	63.324	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 126.640%#	
35) Dibromofluoromethane	8.171	113	85772	57.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 115.180%	
50) Toluene-d8	10.571	98	321895	57.943	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 115.880%#	
62) 4-Bromofluorobenzene	12.853	95	135587	62.603	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 125.200%#	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	24567	37.624	ug/l	90
20) Methylene Chloride	5.271	84	7363	4.900	ug/l #	77
25) 2-Butanone	7.494	43	4563	4.551	ug/l #	90
43) Isopropyl Acetate	8.688	43	221697m	50.725	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
Data File : VN083361.D
Acq On : 16 Aug 2024 18:05
Operator : JC\MD
Sample : PB162767ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162767ZHE#05

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda

Quant Time: Aug 17 01:53:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

