

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
 Data File : VN083363.D
 Acq On : 16 Aug 2024 18:53
 Operator : JC\MD
 Sample : P3629-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-2

Quant Time: Aug 17 01:54:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	253222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107917	61.673	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 123.340%	
35) Dibromofluoromethane	8.171	113	86797	54.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.840%	
50) Toluene-d8	10.571	98	326115	55.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.620%	
62) 4-Bromofluorobenzene	12.847	95	137295	59.731	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 119.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	30222	44.139	ug/l	99
20) Methylene Chloride	5.277	84	8175	5.188	ug/l	84
43) Isopropyl Acetate	8.688	43	400756	90.165	ug/l #	72
52) Toluene	10.629	92	5307	1.179	ug/l	94
95) Naphthalene	15.641	128	22425	2.677	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081624\
Data File : VN083363.D
Acq On : 16 Aug 2024 18:53
Operator : JC\MD
Sample : P3629-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-2

Quant Time: Aug 17 01:54:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

