

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050749.D
 Acq On : 18 Aug 2018 4:09
 Operator : MD\SY
 Sample : J4500-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 18 05:45:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	622105	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1006188	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	873028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309912	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	416214	53.08	ug/l	0.00
Spiked Amount			Recovery	=	106.16%	
35) Dibromofluoromethane	7.59	113	394402	49.10	ug/l	0.00
Spiked Amount			Recovery	=	98.20%	
50) Toluene-d8	10.09	98	1392484	46.06	ug/l	0.00
Spiked Amount			Recovery	=	92.12%	
62) 4-Bromofluorobenzene	12.40	95	381984	38.25	ug/l	0.00
Spiked Amount			Recovery	=	76.50%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	2848	Below Cal		96
5) Bromomethane	2.57	94	3402	Below Cal		95
6) Chloroethane	2.68	64	137	Below Cal	#	43
10) Methyl Iodide	3.95	142	561	7.04	ug/l #	64
11) Tert butyl alcohol	4.80	59	287	0.67	ug/l #	72
14) Allyl chloride	4.55	41	17325	1.65	ug/l #	25
16) Acetone	3.82	43	12865	5.15	ug/l	99
18) Methyl Acetate	4.34	43	12300	1.26	ug/l	95
20) Methylene Chloride	4.55	84	464211	62.60	ug/l	100
25) 2-Butanone	6.85	43	7572	2.36	ug/l	98
30) Chloroform	7.38	83	6512	0.46	ug/l	98
31) Cyclohexane	7.66	56	11204	0.60	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	43578	3.89	ug/l #	48
37) Ethyl Acetate	6.85	43	7572	1.20	ug/l #	76
38) Carbon Tetrachloride	7.67	117	55833	4.81	ug/l #	17
43) Isopropyl Acetate	8.17	43	207225	17.96	ug/l	100
48) Methyl methacrylate	9.18	41	7041	1.21	ug/l #	22
56) Ethyl methacrylate	10.23	69	1463	2.05	ug/l #	53
59) 2-Hexanone	10.77	43	17397	3.45	ug/l #	61
70) Styrene	11.97	104	952	0.98	ug/l #	41
71) Bromoform	12.40	173	2400	0.44	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	12.40	83	148	Below Cal	#	1
76) 1,2,3-Trichloropropane	12.40	75	178991	35.45	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	6320	0.33	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.40	75	178991	128.40	ug/l #	13
84) 1,2,4-Trimethylbenzene	13.04	105	21636	1.11	ug/l	97
95) Naphthalene	15.14	128	42462	7.15	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050749.D
Acq On : 18 Aug 2018 4:09
Operator : MD\SY
Sample : J4500-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 18 05:45:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 622105

Ion Ratio Lower Upper

168 100

99 50.4 40.8 61.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

250000

200000



150000

100000

50000

0

Time--> 7.50 7.60 7.70 7.80



#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 305

Delta R.T. 0.01 min

Lab File: VN050749.D


Acq: 18 Aug 2018 4:09

Tgt Ion: 94 Resp: 3402

Ion Ratio Lower Upper

94 100

96 87.6 74.0 111.0



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

2.57


1500


1000

500

0

Time--> 2.50 2.55 2.60 2.65





#6

Chloroethane

Concen: Below Cal

RT: 2.68 min Scan# 340

Delta R.T. -0.02 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 64 Resp: 137

Ion Ratio Lower Upper




64 100

66 0.0 25.7 38.5#



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0



#10

Methyl Iodide

Concen: 7.04 ug/l

RT: 3.95 min Scan# 733

Delta R.T. -0.01 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09


Tgt Ion: 142 Resp: 561

Ion Ratio Lower Upper

142 100

127 18.2 32.6 49.0#



141 0.0 11.5 17.3#




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11
Tert butyl alcohol
Concen: 0.67 ug/l
RT: 4.80 min Scan# 999
Delta R.T. 0.01 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 59 Resp: 287
Ion Ratio Lower Upper
59 100
57 0.0 8.4 12.6#



#14
Allyl chloride
Concen: 1.65 ug/l
RT: 4.55 min Scan# 921
Delta R.T. 0.23 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09

Tgt Ion: 41 Resp: 17325
Ion Ratio Lower Upper
41 100
39 0.0 51.4 77.0#
76 0.0 29.4 44.0#





#16

Acetone

Concen: 5.15 ug/l

RT: 3.82 min Scan# 695

Delta R.T. 0.01 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 12865

Ion Ratio Lower Upper




43 100

58 33.4 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 1.26 ug/l

RT: 4.34 min Scan# 854

Delta R.T. 0.01 min

Lab File: VN050749.D


Acq: 18 Aug 2018 4:09

Tgt Ion: 43 Resp: 12300

Ion Ratio Lower Upper



43 100


74 22.1 19.7 29.5



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0





#20

Methylene Chloride

Concen: 62.60 ug/l

RT: 4.55 min Scan# 921

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 84 Resp: 464211



Ion Ratio Lower Upper

84 100



49 116.0 92.6 138.8

51 36.3 28.6 43.0

86 64.8 52.2 78.2



Abundance Ion 83.90 (83.60 to 84.60): VNO
Ion 48.90 (48.60 to 49.60): VNO
Ion 50.90 (50.60 to 51.60): VNO
Ion 85.90 (85.60 to 86.60): VNO



#25

2-Butanone

Concen: 2.36 ug/l

RT: 6.85 min Scan# 1636

Delta R.T. 0.01 min

Lab File: VN050749.D



Acq: 18 Aug 2018 4:09

Tgt Ion: 43 Resp: 7572


Ion Ratio Lower Upper


43 100

72 26.4 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 72.00 (71.70 to 72.70): VNO









#30
Chloroform
Concen: 0.46 ug/l
RT: 7.38 min Scan# 1800
Delta R.T. 0.00 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 83 Resp: 6512
Ion Ratio Lower Upper
83 100
85 63.6 52.5 78.7



#31
Cyclohexane
Concen: 0.60 ug/l
RT: 7.66 min Scan# 1888
Delta R.T. 0.01 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09

Tgt Ion: 56 Resp: 11204
Ion Ratio Lower Upper
56 100
69 180.9 25.8 38.6#
84 139.6 67.8 101.6#









#33
 1,2-Dichloroethane-d4
 Concen: 53.08 ug/l
 RT: 8.03 min Scan# 2003
 Delta R.T. 0.00 min
 Lab File: VN050749.D
 Acq: 18 Aug 2018 4:09




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 65 Resp: 416214
 Ion Ratio Lower Upper
 65 100
 67 54.7 0.0 109.8



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2177
 Delta R.T. 0.00 min
 Lab File: VN050749.D
 Acq: 18 Aug 2018 4:09

Tgt Ion: 114 Resp: 1006188
 Ion Ratio Lower Upper
 114 100
 63 19.8 0.0 40.0
 88 15.4 0.0 30.8









#35
Dibromofluoromethane
Concen: 49.10 ug/l
RT: 7.59 min Scan# 1867
Delta R.T. 0.00 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09




Instrument :
MSVOA_N
ClientSampled :


Tot Ion:113 Resp: 394402
Ion Ratio Lower Upper
113 100
111 101.1 81.0 121.6
192 21.5 17.6 26.4



#36
1,1-Dichloropropene
Concen: 3.89 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.13 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09

Tgt Ion: 75 Resp: 43578
Ion Ratio Lower Upper
75 100
110 8.3 18.3 54.9#
77 0.0 25.0 37.4#





#37

Ethyl Acetate

Concen: 1.20 ug/l

RT: 6.85 min Scan# 1636

Delta R.T. -0.08 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 7572

Ion Ratio Lower Upper




43 100

61 6.3 12.0 18.0#

70 0.0 8.5 12.7#



Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 70.00 (69.70 to 70.70): VN0



#38

Carbon Tetrachloride

Concen: 4.81 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.11 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09


Tgt Ion: 117 Resp: 55833

Ion Ratio Lower Upper



117 100


119 6.8 76.6 115.0#

121 0.0 25.0 37.6#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#43

Isopropyl Acetate
 Concen: 17.96 ug/l
 RT: 8.17 min Scan# 2046
 Delta R.T. 0.00 min
 Lab File: VN050749.D
 Acq: 18 Aug 2018 4:09

Instrument :
 MSVOA_N
 ClientSampleId :




Tot Ion: 43 Resp: 207225
 Ion Ratio Lower Upper
 43 100
 61 20.2 16.2 24.2
 87 13.4 10.9 16.3




#48

Methyl methacrylate
 Concen: 1.21 ug/l
 RT: 9.18 min Scan# 2361
 Delta R.T. -0.02 min
 Lab File: VN050749.D
 Acq: 18 Aug 2018 4:09

Tgt Ion: 41 Resp: 7041
 Ion Ratio Lower Upper
 41 100
 69 0.0 73.4 110.0#
 39 19.6 43.0 64.6#









#50
Toluene-d8
Concen: 46.06 ug/l
RT: 10.09 min Scan# 2645
Delta R.T. 0.00 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 98 Resp: 1392484
Ion Ratio Lower Upper
98 100
100 63.7 51.8 77.8



#56
Ethyl methacrylate
Concen: 2.05 ug/l
RT: 10.23 min Scan# 2687
Delta R.T. -0.20 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09

Tgt Ion: 69 Resp: 1463
Ion Ratio Lower Upper
69 100
41 86.8 49.7 74.5#
39 72.1 24.2 36.2#





#59

2-Hexanone

Concen: 3.45 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 17397

Ion Ratio Lower Upper

43 100

58 27.8 28.0 84.0#



#62

4-Bromofluorobenzene

Concen: 38.25 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09




Tgt Ion: 95 Resp: 381984


Ion Ratio Lower Upper

95 100

174 91.1 0.0 177.8

176 87.2 0.0 175.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09




Instrument :
MSVOA_N
ClientSampled :


Tot Ion:117 Resp: 873028
Ion Ratio Lower Upper
117 100
82 54.1 42.4 63.6
119 32.2 25.8 38.8



#70
Styrene
Concen: 0.98 ug/l
RT: 11.97 min Scan# 3227
Delta R.T. 0.00 min
Lab File: VN050749.D
Acq: 18 Aug 2018 4:09

Tgt Ion:104 Resp: 952
Ion Ratio Lower Upper
104 100
78 22.3 39.1 58.7#
103 0.0 44.9 67.3#





#71

Bromoform

Concen: 0.44 ug/l

RT: 12.40 min Scan# 3362

Delta R.T. 0.27 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:173 Resp: 2400

Ion Ratio Lower Upper




173 100

175 892.3 24.4 73.2#

254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09


Tgt Ion:152 Resp: 309912

Ion Ratio Lower Upper



152 100


115 56.3 28.1 84.2

150 158.7 0.0 347.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.40 min Scan# 3363

Delta R.T. -0.10 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 83 Resp: 148

Ion Ratio Lower Upper


83 100

131 347.3 5.3 15.9#



85 0.0 32.1 96.5#



Abundance Ion 82.90 (82.60 to 83.60): VNO
Ion 130.80 (130.50 to 131.50): VNO
Ion 84.90 (84.60 to 85.60): VNO



Time--> 12.38 12.40 12.42



#76

1,2,3-Trichloropropane

Concen: 35.45 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050749.D


Acq: 18 Aug 2018 4:09

Tgt Ion: 75 Resp: 178991


Ion Ratio Lower Upper

75 100


77 1.3 0.0 0.0#




Abundance Ion 74.90 (74.60 to 75.60): VNO
Ion 77.00 (76.70 to 77.70): VNO



Time--> 12.40 12.50





#80

1,3,5-Trimethylbenzene

Concen: 0.33 ug/l

RT: 12.73 min Scan# 3466

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion:105 Resp: 6320

Ion Ratio Lower Upper




105 100

120 49.7 24.7 74.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 128.40 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09


Tgt Ion: 75 Resp: 178991

Ion Ratio Lower Upper

75 100

53 0.2 72.2 108.2#



89 0.0 36.3 54.5#




Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO

Ion 88.90 (88.60 to 89.60): VNO





#84

1,2,4-Trimethylbenzene

Concen: 1.11 ug/l

RT: 13.04 min Scan# 3561

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Instrument :

MSVOA_N





ClientSampled :

Tot Ion:105 Resp: 21636

Ion Ratio Lower Upper

105 100

120 44.4 23.2 69.5



#95

Naphthalene

Concen: 7.15 ug/l

RT: 15.14 min Scan# 4213

Delta R.T. 0.00 min

Lab File: VN050749.D

Acq: 18 Aug 2018 4:09

Tgt Ion:128 Resp: 42462

Ion Ratio Lower Upper

128 100

127 13.2 10.3 15.5

129 10.6 8.5 12.7

