

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050716.D
 Acq On : 17 Aug 2018 8:43
 Operator : MD\SY
 Sample : VN0817WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0817WBL01

Quant Time: Aug 17 14:27:45 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA N\methods\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	117241	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	584641	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	477796	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	231095	29.00	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.67%
60) 4-Bromofluorobenzene	12.40	95	156295	21.29	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	70.97%
63) Toluene-d8	10.09	98	691078	30.28	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.93%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050716.D
Acq On : 17 Aug 2018 8:43
Operator : MD\SY
Sample : VN0817WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0817WBL01

Quant Time: Aug 17 14:27:45 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA N\methods\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

