

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050718.D
 Acq On : 17 Aug 2018 9:37
 Operator : MD\SY
 Sample : J4527-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Aug 17 14:32:40 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA N\methods\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	113979	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	568052	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	469179	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	226882	29.29	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	97.63%		
60) 4-Bromofluorobenzene	12.40	95	152822	21.20	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	70.67%		
63) Toluene-d8	10.09	98	661109	29.50	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	98.33%		

Target Compounds					Ovalue
25) Chloroform	7.37	83	22490	1.80 ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050718.D
Acq On : 17 Aug 2018 9:37
Operator : MD\SY
Sample : J4527-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Aug 17 14:32:40 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA N\methods\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

