

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050722.D
 Acq On : 17 Aug 2018 11:55
 Operator : MD\SY
 Sample : VN0817WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0817WBL02

Quant Time: Aug 18 00:05:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	669682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1049802	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	896716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	301897	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	437153	51.79	ug/l	0.00
Spiked Amount			Recovery	=	103.58%	
35) Dibromofluoromethane	7.59	113	415124	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
50) Toluene-d8	10.09	98	1479592	46.91	ug/l	0.00
Spiked Amount			Recovery	=	93.82%	
62) 4-Bromofluorobenzene	12.40	95	384596	36.91	ug/l	0.00
Spiked Amount			Recovery	=	73.82%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050722.D
Acq On : 17 Aug 2018 11:55
Operator : MD\SY
Sample : VN0817WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0817WBL02

Quant Time: Aug 18 00:05:46 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

