

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050726.D
 Acq On : 17 Aug 2018 13:33
 Operator : MD\SY
 Sample : PB112140TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112140TB

Quant Time: Aug 18 00:16:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	624698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1006588	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	863112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282680	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	421142	53.48	ug/l	0.00
Spiked Amount						
			Recovery	=	106.96%	
35) Dibromofluoromethane	7.59	113	393791	49.00	ug/l	0.00
Spiked Amount						
			Recovery	=	98.00%	
50) Toluene-d8	10.09	98	1402247	46.36	ug/l	0.00
Spiked Amount						
			Recovery	=	92.72%	
62) 4-Bromofluorobenzene	12.40	95	367344	36.76	ug/l	0.00
Spiked Amount						
			Recovery	=	73.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	27298	12.51	ug/l	97
18) Methyl Acetate	4.33	43	11615	1.13	ug/l #	82
20) Methylene Chloride	4.56	84	20449	1.50	ug/l	96
43) Isopropyl Acetate	8.17	43	240348	20.82	ug/l	99
70) Styrene	11.97	104	1574	1.01	ug/l #	85
95) Naphthalene	15.14	128	7597	4.51	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050726.D
Acq On : 17 Aug 2018 13:33
Operator : MD\SY
Sample : PB112140TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112140TB

Quant Time: Aug 18 00:16:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

