

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050728.D
 Acq On : 17 Aug 2018 14:23
 Operator : MD\SY
 Sample : PB112140ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112140ZHE#02

Quant Time: Aug 18 00:21:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	617542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1024461	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	853199	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	269630	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	418060	53.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.42%	
35) Dibromofluoromethane	7.59	113	390096	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
50) Toluene-d8	10.09	98	1382977	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
62) 4-Bromofluorobenzene	12.40	95	358294	35.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.46%	
Target Compounds						
16) Acetone	3.83	43	26166	12.08	ug/l	95
18) Methyl Acetate	4.33	43	12403	1.30	ug/l	100
20) Methylene Chloride	4.55	84	20727	1.57	ug/l	98
43) Isopropyl Acetate	8.17	43	262908	22.38	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050728.D
Acq On : 17 Aug 2018 14:23
Operator : MD\SY
Sample : PB112140ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112140ZHE#02

Quant Time: Aug 18 00:21:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

