

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050730.D
 Acq On : 17 Aug 2018 15:13
 Operator : MD\SY
 Sample : PB112140ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112140ZHE#04

Quant Time: Aug 18 00:27:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	629844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1061318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	908422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	320448	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	431888	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
35) Dibromofluoromethane	7.59	113	404015	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	10.09	98	1442867	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
62) 4-Bromofluorobenzene	12.40	95	400331	38.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.00%	
Target Compounds						
16) Acetone	3.82	43	26185	11.83	ug/l	92
18) Methyl Acetate	4.33	43	16128	1.94	ug/l	93
20) Methylene Chloride	4.55	84	21134	1.57	ug/l #	91
43) Isopropyl Acetate	8.17	43	283556	23.30	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050730.D
Acq On : 17 Aug 2018 15:13
Operator : MD\SY
Sample : PB112140ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112140ZHE#04

Quant Time: Aug 18 00:27:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

