

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050738.D
 Acq On : 17 Aug 2018 18:31
 Operator : MD\SY
 Sample : PB112140ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112140ZHE#05

Quant Time: Aug 18 01:17:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	591006	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1011610	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	871236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	319113	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	416369	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
35) Dibromofluoromethane	7.59	113	387590	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	10.09	98	1369712	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
62) 4-Bromofluorobenzene	12.40	95	383154	38.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.32%	
Target Compounds						
16) Acetone	3.82	43	26958	13.12	ug/l	92
18) Methyl Acetate	4.34	43	17793	2.45	ug/l #	73
20) Methylene Chloride	4.56	84	21092	1.75	ug/l	97
43) Isopropyl Acetate	8.17	43	304332	26.23	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050738.D
Acq On : 17 Aug 2018 18:31
Operator : MD\SY
Sample : PB112140ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112140ZHE#05

Quant Time: Aug 18 01:17:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

