

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050747.D
 Acq On : 18 Aug 2018 3:19
 Operator : MD\SY
 Sample : J4486-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ202

Manual Integrations
 APPROVED

MMDadoda
 8/20/2018 10:53:43 AM

Quant Time: Aug 18 05:39:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	648587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1030345	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	869635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	284662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	424732	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
35) Dibromofluoromethane	7.59	113	400795	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
50) Toluene-d8	10.09	98	1434921	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
62) 4-Bromofluorobenzene	12.40	95	375666	36.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.46%	
Target Compounds						
16) Acetone	3.82	43	29740	13.20	ug/l	98
18) Methyl Acetate	4.34	43	13005	1.30	ug/l #	88
20) Methylene Chloride	4.53	84	38062740m	5024.78	ug/l	
28) Bromochloromethane	7.20	49	91623	13.86	ug/l	99
43) Isopropyl Acetate	8.17	43	224090	18.97	ug/l	100
95) Naphthalene	15.13	128	15060	5.13	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050747.D
Acq On : 18 Aug 2018 3:19
Operator : MD\SY
Sample : J4486-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ202

Manual Integrations
APPROVED

MMDadoda
8/20/2018 10:53:43 AM

Quant Time: Aug 18 05:39:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

