

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050755.D
 Acq On : 18 Aug 2018 6:37
 Operator : MD\SY
 Sample : J4521-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-12003

Quant Time: Aug 20 00:59:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	612207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1014324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	862458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	300242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	418651	54.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.50%	
35) Dibromofluoromethane	7.59	113	391586	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	10.09	98	1391445	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
62) 4-Bromofluorobenzene	12.40	95	368806	36.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.26%	
Target Compounds						
16) Acetone	3.83	43	13030	5.34	ug/l #	86
18) Methyl Acetate	4.34	43	15840	1.97	ug/l #	89
20) Methylene Chloride	4.55	84	183455	24.36	ug/l	98
30) Chloroform	7.37	83	16047	1.16	ug/l	96
43) Isopropyl Acetate	8.17	43	229793	19.76	ug/l	98
95) Naphthalene	15.14	128	120228	13.47	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050755.D
Acq On : 18 Aug 2018 6:37
Operator : MD\SY
Sample : J4521-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-12003

Quant Time: Aug 20 00:59:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

