

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
 Data File : VN050759.D
 Acq On : 18 Aug 2018 8:16
 Operator : MD\SY
 Sample : J4521-27
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11988

Quant Time: Aug 20 01:11:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	651481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1060710	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	882924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	271341	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	432850	52.71	ug/l	0.00
Spiked Amount			Recovery	=	105.42%	
35) Dibromofluoromethane	7.59	113	411545	48.60	ug/l	0.00
Spiked Amount			Recovery	=	97.20%	
50) Toluene-d8	10.09	98	1465631	45.99	ug/l	0.00
Spiked Amount			Recovery	=	91.98%	
62) 4-Bromofluorobenzene	12.40	95	367646	34.92	ug/l	0.00
Spiked Amount			Recovery	=	69.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	14049	5.43	ug/l	95
18) Methyl Acetate	4.33	43	15931	1.80	ug/l	95
20) Methylene Chloride	4.55	84	40108	3.97	ug/l	97
30) Chloroform	7.37	83	28703	1.94	ug/l	94
43) Isopropyl Acetate	8.17	43	254872	20.95	ug/l	96
95) Naphthalene	15.14	128	33809	6.85	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081718\
Data File : VN050759.D
Acq On : 18 Aug 2018 8:16
Operator : MD\SY
Sample : J4521-27
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11988

Quant Time: Aug 20 01:11:58 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

