

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081719\
 Data File : VN057353.D
 Acq On : 16 Aug 2019 14:18
 Operator : JC/SP
 Sample : VN0817WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0817WBL02

Quant Time: Aug 16 14:40:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 16 13:33:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	142114	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	781338	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	600895	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.02	65	278819	30.52	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.73%
60) 4-Bromofluorobenzene	12.40	95	192200	21.72	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	72.40%
63) Toluene-d8	10.08	98	894761	33.01	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	110.03%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081719\
Data File : VN057353.D
Acq On : 16 Aug 2019 14:18
Operator : JC/SP
Sample : VN0817WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0817WBL02

Quant Time: Aug 16 14:40:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N081719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 16 13:33:03 2019
Response via : Initial Calibration

