

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081720\
 Data File : VN062983.D
 Acq On : 17 Aug 2020 21:08
 Operator : JC/MD
 Sample : PB130989TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130989TB

Quant Time: Aug 18 07:26:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 17 17:31:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	185665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	286349	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	295712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	114298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102098	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
35) Dibromofluoromethane	7.55	113	97708	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
50) Toluene-d8	10.06	98	359253	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	123785	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
Target Compounds						
16) Acetone	3.78	43	3875	5.088	ug/l	91
20) Methylene Chloride	4.50	84	3373	2.422	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081720\
Data File : VN062983.D
Acq On : 17 Aug 2020 21:08
Operator : JC/MD
Sample : PB130989TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130989TB

Quant Time: Aug 18 07:26:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 17 17:31:03 2020
Response via : Initial Calibration

