

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081720\
 Data File : VN062984.D
 Acq On : 17 Aug 2020 21:32
 Operator : JC/MD
 Sample : PB130988TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130988TB

Quant Time: Aug 18 07:29:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 17 17:31:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	197443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	299738	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	325758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	121134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106803	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
35) Dibromofluoromethane	7.55	113	109454	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
50) Toluene-d8	10.06	98	391329	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	12.38	95	134062	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
Target Compounds						
16) Acetone	3.78	43	10530	13.002	ug/l	96
20) Methylene Chloride	4.50	84	9728	5.521	ug/l	94
43) Isopropyl Acetate	8.13	43	43948	11.776	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081720\
Data File : VN062984.D
Acq On : 17 Aug 2020 21:32
Operator : JC/MD
Sample : PB130988TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130988TB

Quant Time: Aug 18 07:29:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 17 17:31:03 2020
Response via : Initial Calibration

