

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081720\
 Data File : VN062990.D
 Acq On : 17 Aug 2020 23:57
 Operator : JC/MD
 Sample : L3696-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 W-1

Quant Time: Aug 18 07:46:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 17 17:31:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	187720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	302942	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	315590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	132982	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108181	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
35) Dibromofluoromethane	7.55	113	98906	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
50) Toluene-d8	10.06	98	380070	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	134828	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
Target Compounds						
					Qvalue	
16) Acetone	3.79	43	5186	6.735	ug/l	99
73) Isopropylbenzene	12.23	105	6946	0.757	ug/l	93
78) n-propylbenzene	12.57	91	11060	1.033	ug/l	95
84) 1,2,4-Trimethylbenzene	13.02	105	14247	1.907	ug/l	98
85) sec-Butylbenzene	13.15	105	7413	0.846	ug/l	97
86) p-Isopropyltoluene	13.26	119	2885	0.370	ug/l	93
89) n-Butylbenzene	13.59	91	3516	0.555	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081720\
Data File : VN062990.D
Acq On : 17 Aug 2020 23:57
Operator : JC/MD
Sample : L3696-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
W-1

Quant Time: Aug 18 07:46:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 17 17:31:03 2020
Response via : Initial Calibration

