


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081720\
 Data File : VN062967.D
 Acq On : 17 Aug 2020 12:22
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081720W.M
 Title : SW846 8260
 Last Update : Mon Aug 17 17:31:03 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	6014	PASS
75	95	30	60	50.1	17121	PASS
95	95	100	100	100.0	34178	PASS
96	95	5	9	7.4	2514	PASS
173	174	0.00	2	1.7	516	PASS
174	95	50	100	90.7	30989	PASS
175	174	5	9	7.1	2211	PASS
176	174	95	101	99.3	30778	PASS
177	176	5	9	6.8	2107	PASS