

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068145.D
 Acq On : 17 Aug 2021 16:04
 Operator : JC/MD
 Sample : M3408-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33S

Quant Time: Aug 17 16:30:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	444257	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	731087	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	775624	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	293504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	246873	52.045	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.100%	
35) Dibromofluoromethane	8.024	113	224579	48.928	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.860%	
50) Toluene-d8	10.443	98	920006	52.604	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	12.733	95	296579	50.940	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	22395	16.170	ug/l	# 88
20) Methylene Chloride	5.095	84	27867	5.375	ug/l	# 78
25) 2-Butanone	7.353	43	5357	2.482	ug/l	97
43) Isopropyl Acetate	8.563	43	34963	4.287	ug/l	# 93
93) 1,2,4-Trichlorobenzene	15.268	180	7907	4.213	ug/l	97
95) Naphthalene	15.507	128	29548	5.737	ug/l	98
96) 1,2,3-Trichlorobenzene	15.708	180	9322	4.723	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068145.D
Acq On : 17 Aug 2021 16:04
Operator : JC/MD
Sample : M3408-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33S

Quant Time: Aug 17 16:30:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

