

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068151.D
 Acq On : 17 Aug 2021 18:31
 Operator : JC/MD
 Sample : M3361-12 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41940

Quant Time: Aug 18 07:20:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	406941	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	678786	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	662551	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	251242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	233933	53.839	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.680%	
35) Dibromofluoromethane	8.024	113	210102	49.301	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.600%	
50) Toluene-d8	10.443	98	830007	51.115	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.220%	
62) 4-Bromofluorobenzene	12.733	95	265867	49.183	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.360%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	43067	33.948	ug/l	92
40) Benzene	8.464	78	4782	0.282	ug/l	90
52) Toluene	10.505	92	9481	0.821	ug/l	99
95) Naphthalene	15.509	128	5599	3.874	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068151.D
Acq On : 17 Aug 2021 18:31
Operator : JC/MD
Sample : M3361-12 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41940

Quant Time: Aug 18 07:20:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

