

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068160.D
 Acq On : 17 Aug 2021 23:25
 Operator : JC/MD
 Sample : VN0817WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0817WBL02

Quant Time: Aug 18 07:22:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	426481	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	693639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	696691	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	280959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232711	51.104	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.200%	
35) Dibromofluoromethane	8.024	113	209931	48.206	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.420%	
50) Toluene-d8	10.443	98	840947	50.679	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	12.734	95	281756	51.007	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.020%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068160.D
Acq On : 17 Aug 2021 23:25
Operator : JC/MD
Sample : VN0817WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0817WBL02

Quant Time: Aug 18 07:22:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

