

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068165.D
 Acq On : 18 Aug 2021 1:28
 Operator : JC/MD
 Sample : M3435-33
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-11-EB-02

Quant Time: Aug 18 08:40:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	390010	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	645762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	647331	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	270893	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222869	53.520	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	107.040%	
35) Dibromofluoromethane	8.024	113	202550	49.960	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	99.920%	
50) Toluene-d8	10.446	98	784050	50.754	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.500%	
62) 4-Bromofluorobenzene	12.734	95	267407	51.998	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	104.000%	
Target Compounds						
						Qvalue
16) Acetone	4.291	43	74654	61.402	ug/l	# 88
25) 2-Butanone	7.345	43	9784	5.164	ug/l	# 75
95) Naphthalene	15.512	128	6649	3.928	ug/l	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068165.D
Acq On : 18 Aug 2021 1:28
Operator : JC/MD
Sample : M3435-33
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-11-EB-02

Quant Time: Aug 18 08:40:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

