

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068167.D
 Acq On : 18 Aug 2021 2:17
 Operator : JC/MD
 Sample : M3435-37
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-13-EB-04

Quant Time: Aug 18 08:41:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	409320	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	674202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	685036	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	265344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231768	53.031	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.060%	
35) Dibromofluoromethane	8.024	113	210377	49.701	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.400%	
50) Toluene-d8	10.443	98	827774	51.324	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.640%	
62) 4-Bromofluorobenzene	12.733	95	272018	50.664	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.320%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	72168	56.557	ug/l	91
25) 2-Butanone	7.348	43	9307	4.680	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068167.D
Acq On : 18 Aug 2021 2:17
Operator : JC/MD
Sample : M3435-37
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-13-EB-04

Quant Time: Aug 18 08:41:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

