

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068170.D
 Acq On : 18 Aug 2021 3:30
 Operator : JC/MD
 Sample : M3446-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW6B-GW-20210816

Quant Time: Aug 18 08:41:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	415890	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	694776	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	690793	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	260250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237416	53.465	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	106.940%		
35) Dibromofluoromethane	8.024	113	214494	49.173	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.340%		
50) Toluene-d8	10.443	98	846952	50.958	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.920%		
62) 4-Bromofluorobenzene	12.736	95	273574	49.444	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.880%		
Target Compounds						
						Qvalue
3) Chloromethane	2.303	50	4015	1.078	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.208	101	28309	8.073	ug/l #	69
12) 1,1-Dichloroethene	4.181	96	7948	2.357	ug/l #	71
27) cis-1,2-Dichloroethene	7.329	96	4970	1.100	ug/l	59
30) Chloroform	7.817	83	10429	1.392	ug/l	99
38) Carbon Tetrachloride	8.209	117	20245	3.091	ug/l	97
44) Trichloroethene	9.217	130	1101764	211.353	ug/l	78
64) Tetrachloroethene	10.977	164	1432	0.268	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068170.D
Acq On : 18 Aug 2021 3:30
Operator : JC/MD
Sample : M3446-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW6B-GW-20210816

Quant Time: Aug 18 08:41:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

