

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068171.D
 Acq On : 18 Aug 2021 3:55
 Operator : JC/MD
 Sample : M3410-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

MMDadoda
 8/19/2021 5:54:57 PM

Quant Time: Aug 18 08:41:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	393806	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	659080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	650213	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	240562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230543	54.829	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 109.660%		
35) Dibromofluoromethane	8.024	113	205514	49.666	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 99.340%		
50) Toluene-d8	10.446	98	792098	50.238	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 100.480%		
62) 4-Bromofluorobenzene	12.736	95	253262	48.252	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 96.500%		
Target Compounds						Qvalue
16) Acetone	4.293	43	4232m	3.447	ug/l	
19) Methyl tert-butyl Ether	5.624	73	11337m	1.101	ug/l	
44) Trichloroethene	9.223	130	4540	0.918	ug/l	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068171.D
Acq On : 18 Aug 2021 3:55
Operator : JC/MD
Sample : M3410-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Manual Integrations
APPROVED

MMDadoda
8/19/2021 5:54:57 PM

Quant Time: Aug 18 08:41:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

