

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068172.D
 Acq On : 18 Aug 2021 4:20
 Operator : JC/MD
 Sample : M3410-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Aug 18 08:41:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	394569	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	659558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	651803	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	245384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229731	54.530	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	109.060%	
35) Dibromofluoromethane	8.024	113	203533	49.152	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	98.300%	
50) Toluene-d8	10.446	98	797571	50.549	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.100%	
62) 4-Bromofluorobenzene	12.736	95	257856	49.092	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	98.180%	
Target Compounds						
						Qvalue
16) Acetone	4.304	43	3470	2.821	ug/l	# 75
44) Trichloroethene	9.220	130	2047	0.414	ug/l	77
64) Tetrachloroethene	10.979	164	10313	2.046	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068172.D
Acq On : 18 Aug 2021 4:20
Operator : JC/MD
Sample : M3410-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Aug 18 08:41:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

