

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068175.D
 Acq On : 18 Aug 2021 5:33
 Operator : JC/MD
 Sample : M3410-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Aug 18 08:42:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	415664	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	681878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	673034	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	258046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229571	51.727	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.460%
35) Dibromofluoromethane	8.024	113	211383	49.377	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.760%
50) Toluene-d8	10.446	98	823403	50.478	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.733	95	264852	48.774	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.540%
Target Compounds						
						Qvalue
4) Vinyl Chloride	2.443	62	38116	6.537	ug/l	100
27) cis-1,2-Dichloroethene	7.329	96	253532	56.126	ug/l	68
44) Trichloroethene	9.217	130	36000	7.037	ug/l	80
64) Tetrachloroethene	10.979	164	76792	14.756	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068175.D
Acq On : 18 Aug 2021 5:33
Operator : JC/MD
Sample : M3410-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Aug 18 08:42:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

