

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068177.D
 Acq On : 18 Aug 2021 6:22
 Operator : JC/MD
 Sample : M3432-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 18 08:42:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	387686	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	650487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	654541	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	243478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222204	53.680	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.360%
35) Dibromofluoromethane	8.024	113	201585	49.361	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.720%
50) Toluene-d8	10.443	98	790793	50.818	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.640%
62) 4-Bromofluorobenzene	12.734	95	252815	48.804	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068177.D
Acq On : 18 Aug 2021 6:22
Operator : JC/MD
Sample : M3432-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 18 08:42:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

