

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068159.D
 Acq On : 17 Aug 2021 22:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 18 07:22:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.396	0.405	-2.3	92	0.00
3 P	Chloromethane	0.448	0.442	1.3	94	0.00
4 C	Vinyl Chloride	0.701	0.727	-3.7#	98	0.00
5 T	Bromomethane	0.636	0.642	-0.9	101	0.00
6 T	Chloroethane	0.535	0.540	-0.9	98	0.00
7 T	Trichlorofluoromethane	1.379	1.417	-2.8	98	0.00
8 T	Diethyl Ether	0.320	0.318	0.6	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.422	0.443	-5.0	98	0.00
10 T	Methyl Iodide	0.599	0.663	-10.7	94	0.00
11 T	Tert butyl alcohol	0.067	0.072	-7.5	100	0.00
12 CM	1,1-Dichloroethene	0.405	0.421	-4.0#	98	0.00
13 T	Acrolein	0.050	0.050	0.0	103	0.00
14 T	Allyl chloride	0.518	0.541	-4.4	96	0.00
15 T	Acrylonitrile	0.188	0.209	-11.2	98	0.00
16 T	Acetone	0.156	0.158	-1.3	98	0.00
17 T	Carbon Disulfide	1.048	1.121	-7.0	97	0.00
18 T	Methyl Acetate	0.567	0.467	17.6	98	0.00
19 T	Methyl tert-butyl Ether	1.307	1.452	-11.1	96	0.00
20 T	Methylene Chloride	0.524	0.498	5.0	98	0.00
21 T	trans-1,2-Dichloroethene	0.454	0.478	-5.3	95	0.00
22 T	Diisopropyl ether	1.132	1.288	-13.8	97	0.00
23 T	Vinyl Acetate	0.881	1.006	-14.2	96	0.00
24 P	1,1-Dichloroethane	0.756	0.795	-5.2	95	0.00
25 T	2-Butanone	0.243	0.265	-9.1	97	0.00
26 T	2,2-Dichloropropane	0.648	0.636	1.9	87	0.00
27 T	cis-1,2-Dichloroethene	0.543	0.578	-6.4	96	0.00
28 T	Bromochloromethane	0.333	0.323	3.0	88	0.00
29 T	Tetrahydrofuran	0.155	0.177	-14.2	98	0.00
30 C	Chloroform	0.901	0.965	-7.1#	97	0.00
31 T	Cyclohexane	0.703	0.718	-2.1	96	0.00
32 T	1,1,1-Trichloroethane	0.803	0.894	-11.3	97	0.00
33 S	1,2-Dichloroethane-d4	0.534	0.554	-3.7	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.314	0.335	-6.7	96	0.00
36 T	1,1-Dichloropropene	0.409	0.464	-13.4	97	0.00
37 T	Ethyl Acetate	0.348	0.371	-6.6	96	0.00
38 T	Carbon Tetrachloride	0.471	0.545	-15.7	98	0.00
39 T	Methylcyclohexane	0.462	0.528	-14.3	95	0.00
40 TM	Benzene	1.248	1.411	-13.1	96	0.00
41 T	Methacrylonitrile	0.159	0.187	-17.6	102	0.00
42 TM	1,2-Dichloroethane	0.425	0.476	-12.0	98	0.00
43 T	Isopropyl Acetate	0.558	0.641	-14.9	97	0.00
44 TM	Trichloroethene	0.375	0.419	-11.7	95	0.00
45 C	1,2-Dichloropropane	0.293	0.333	-13.7#	97	0.00
46 T	Dibromomethane	0.239	0.268	-12.1	97	0.00
47 T	Bromodichloromethane	0.445	0.515	-15.7	98	0.00
48 T	Methyl methacrylate	0.232	0.276	-19.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
 Data File : VN068159.D
 Acq On : 17 Aug 2021 22:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 18 07:22:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	99	0.00
50 S	Toluene-d8	1.196	1.338	-11.9	98	0.00
51 T	4-Methyl-2-Pentanone	0.343	0.409	-19.2	98	0.00
52 CM	Toluene	0.850	1.007	-18.5#	98	0.00
53 T	t-1,3-Dichloropropene	0.437	0.535	-22.4	97	0.00
54 T	cis-1,3-Dichloropropene	0.473	0.552	-16.7	96	0.00
55 T	1,1,2-Trichloroethane	0.340	0.395	-16.2	99	0.00
56 T	Ethyl methacrylate	0.436	0.540	-23.9	98	0.00
57 T	1,3-Dichloropropane	0.523	0.598	-14.3	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.072	0.091	-26.4#	96	0.00
59 T	2-Hexanone	0.238	0.289	-21.4	97	0.00
60 T	Dibromochloromethane	0.383	0.470	-22.7	98	0.00
61 T	1,2-Dibromoethane	0.352	0.423	-20.2	98	0.00
62 S	4-Bromofluorobenzene	0.398	0.474	-19.1	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.387	0.444	-14.7	98	0.00
65 PM	Chlorobenzene	1.024	1.141	-11.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.456	-18.1	99	0.00
67 C	Ethyl Benzene	1.689	1.966	-16.4#	98	0.00
68 T	m/p-Xylenes	0.676	0.808	-19.5	98	0.00
69 T	o-Xylene	0.653	0.789	-20.8	99	0.00
70 T	Styrene	1.045	1.301	-24.5	98	0.00
71 P	Bromoform	0.310	0.397	-28.1#	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.556	3.817	-7.3	98	0.00
74 T	N-amyl acetate	0.926	0.995	-7.5	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.066	1.066	0.0	98	0.00
76 T	1,2,3-Trichloropropane	0.831	0.772	7.1	89	0.00
77 T	Bromobenzene	0.980	1.083	-10.5	98	0.00
78 T	n-propylbenzene	3.815	4.172	-9.4	98	0.00
79 T	2-Chlorotoluene	2.359	2.509	-6.4	99	0.00
80 T	1,3,5-Trimethylbenzene	2.877	3.181	-10.6	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.254	0.279	-9.8	91	0.00
82 T	4-Chlorotoluene	2.290	2.475	-8.1	98	0.00
83 T	tert-Butylbenzene	2.555	2.809	-9.9	98	0.00
84 T	1,2,4-Trimethylbenzene	2.808	3.140	-11.8	98	0.00
85 T	sec-Butylbenzene	3.364	3.702	-10.0	98	0.00
86 T	p-Isopropyltoluene	2.776	3.172	-14.3	97	0.00
87 T	1,3-Dichlorobenzene	1.671	1.902	-13.8	97	0.00
88 T	1,4-Dichlorobenzene	1.692	1.861	-10.0	97	0.00
89 T	n-Butylbenzene	2.071	2.284	-10.3	96	0.00
90 T	Hexachloroethane	0.515	0.567	-10.1	97	0.00
91 T	1,2-Dichlorobenzene	1.591	1.798	-13.0	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.162	0.172	-6.2	96	0.00
93 T	1,2,4-Trichlorobenzene	0.693	0.826	-19.2	92	0.00
94 T	Hexachlorobutadiene	0.431	0.476	-10.4	96	0.00
95 T	Naphthalene	1.486	1.913	-28.7#	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081721\
Data File : VN068159.D
Acq On : 17 Aug 2021 22:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 18 07:22:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.638	0.784	-22.9	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6