

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN081722\  
 Data File : VN073986.D  
 Acq On : 17 Aug 2022 19:14  
 Operator : JC\MD  
 Sample : PB147075TB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB147075TB

Quant Time: Aug 18 00:49:49 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N081522W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 16 00:45:21 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.016	168	643316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	940741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	932139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	474590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	307829	49.128	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.260%
35) Dibromofluoromethane	7.951	113	298585	53.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.340%
50) Toluene-d8	10.381	98	1021186	47.394	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.780%
62) 4-Bromofluorobenzene	12.675	95	429184	60.018	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	120.040%
Target Compounds						
						Qvalue
16) Acetone	4.228	43	100125	33.870	ug/l	99
20) Methylene Chloride	5.010	84	21148	3.390	ug/l	98
37) Ethyl Acetate	7.340	43	81322	8.687	ug/l	99
43) Isopropyl Acetate	8.498	43	342846	25.707	ug/l #	92
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN081722\  
 Data File : VN073986.D  
 Acq On : 17 Aug 2022 19:14  
 Operator : JC\MD  
 Sample : PB147075TB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB147075TB

Quant Time: Aug 18 00:49:49 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N081522W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 16 00:45:21 2022  
 Response via : Initial Calibration

