

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081819\
 Data File : VN057378.D
 Acq On : 17 Aug 2019 1:15
 Operator : JC/SP
 Sample : K3616-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-081519-B

Quant Time: Aug 17 04:36:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1067002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1495670	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1256639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	348018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	504037	43.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.90%	
35) Dibromofluoromethane	7.58	113	454809	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
50) Toluene-d8	10.09	98	1740850	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	12.40	95	476583	37.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.22%	
Target Compounds						
20) Methylene Chloride	4.54	84	49970	4.438	ug/l	96
43) Isopropyl Acetate	8.16	43	329663	15.897	ug/l #	73
95) Naphthalene	15.13	128	23579	2.023	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081819\
Data File : VN057378.D
Acq On : 17 Aug 2019 1:15
Operator : JC/SP
Sample : K3616-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-081519-B

Quant Time: Aug 17 04:36:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

