

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068188.D
 Acq On : 18 Aug 2021 13:06
 Operator : JC/MD
 Sample : M3361-14DL 40X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 41909DL

Quant Time: Aug 18 13:44:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	424833	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	683984	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	700502	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	290807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237300	52.314	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 104.620%		
35) Dibromofluoromethane	8.021	113	212125	49.398	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 98.800%		
50) Toluene-d8	10.446	98	833827	50.960	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 101.920%		
62) 4-Bromofluorobenzene	12.733	95	286507	52.599	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 105.200%		
Target Compounds						
						Qvalue
16) Acetone	4.299	43	22552	17.028	ug/l	# 82
19) Methyl tert-butyl Ether	5.618	73	16874	1.520	ug/l	98
25) 2-Butanone	7.345	43	117256	56.811	ug/l	# 89
37) Ethyl Acetate	7.420	43	143808	30.191	ug/l	# 94
40) Benzene	8.461	78	130303	7.635	ug/l	97
52) Toluene	10.507	92	423717	36.420	ug/l	99
67) Ethyl Benzene	11.846	91	67951	2.872	ug/l	97
68) m/p-Xylenes	11.953	106	94714	10.008	ug/l	81
69) o-Xylene	12.283	106	57457	6.283	ug/l	80
78) n-propylbenzene	12.924	91	11636	0.524	ug/l	86
80) 1,3,5-Trimethylbenzene	13.061	105	19028	1.137	ug/l	97
84) 1,2,4-Trimethylbenzene	13.372	105	91507	5.602	ug/l	91
95) Naphthalene	15.509	128	45935	7.099	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068188.D
Acq On : 18 Aug 2021 13:06
Operator : JC/MD
Sample : M3361-14DL 40X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41909DL

Quant Time: Aug 18 13:44:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

