

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068191.D
 Acq On : 18 Aug 2021 14:21
 Operator : JC/MD
 Sample : M3446-02DL 4X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW6B-GW-20210816DL

Manual Integrations
 APPROVED

MMDadoda
 8/19/2021 6:17:29 PM

Quant Time: Aug 19 03:06:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	416421	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	679166	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	693029	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	262272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232985	52.400	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.800%
35) Dibromofluoromethane	8.024	113	214540	50.314	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.620%
50) Toluene-d8	10.446	98	825918	50.834	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.660%
62) 4-Bromofluorobenzene	12.736	95	266963	49.359	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.720%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.191	101	6556m	1.867	ug/l	
30) Chloroform	7.815	83	2808	0.374	ug/l	100
38) Carbon Tetrachloride	8.201	117	4912	0.767	ug/l	# 78
44) Trichloroethene	9.217	130	247803	48.629	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068191.D
Acq On : 18 Aug 2021 14:21
Operator : JC/MD
Sample : M3446-02DL 4X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW6B-GW-20210816DL

Manual Integrations
APPROVED

MMDadoda
8/19/2021 6:17:29 PM

Quant Time: Aug 19 03:06:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

