

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068193.D
 Acq On : 18 Aug 2021 15:10
 Operator : JC/MD
 Sample : M3435-31
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-10-EB-01

Quant Time: Aug 19 03:07:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	388387	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	647832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	658785	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	251301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225244	54.316	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.640%	
35) Dibromofluoromethane	8.024	113	202346	49.750	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.500%	
50) Toluene-d8	10.446	98	796346	51.385	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.760%	
62) 4-Bromofluorobenzene	12.733	95	257994	50.007	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.020%	
Target Compounds						
16) Acetone	4.293	43	4452	3.677	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068193.D
Acq On : 18 Aug 2021 15:10
Operator : JC/MD
Sample : M3435-31
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-10-EB-01

Quant Time: Aug 19 03:07:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

