

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068197.D
 Acq On : 18 Aug 2021 16:50
 Operator : JC/MD
 Sample : M3463-41
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-17-EB

Quant Time: Aug 19 03:07:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	400710	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	669007	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	672016	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	259679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	228751	53.465	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 106.940%		
35) Dibromofluoromethane	8.024	113	211026	50.242	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.480%		
50) Toluene-d8	10.446	98	811610	50.712	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 101.420%		
62) 4-Bromofluorobenzene	12.734	95	271494	50.959	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 101.920%		
Target Compounds						
						Qvalue
16) Acetone	4.296	43	27635	22.122	ug/l	# 90
25) 2-Butanone	7.351	43	12550	6.447	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068197.D
Acq On : 18 Aug 2021 16:50
Operator : JC/MD
Sample : M3463-41
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-17-EB

Quant Time: Aug 19 03:07:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

