

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068199.D
 Acq On : 18 Aug 2021 17:39
 Operator : JC/MD
 Sample : M3435-37
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-13-EB-04

Quant Time: Aug 19 03:07:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	385846	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	638163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	647619	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	247030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	220591	53.544	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.080%	
35) Dibromofluoromethane	8.024	113	201467	50.284	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.560%	
50) Toluene-d8	10.443	98	783414	51.316	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.640%	
62) 4-Bromofluorobenzene	12.736	95	258948	50.953	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.900%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	74318	61.785	ug/l	94
25) 2-Butanone	7.351	43	8911	4.754	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068199.D
Acq On : 18 Aug 2021 17:39
Operator : JC/MD
Sample : M3435-37
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-13-EB-04

Quant Time: Aug 19 03:07:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

