

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
 Data File : VN068205.D
 Acq On : 18 Aug 2021 20:08
 Operator : JC/MD
 Sample : M3449-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1

Quant Time: Aug 19 03:08:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	418951	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	734971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	786350	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	301042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	245859	54.962	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.920%	
35) Dibromofluoromethane	8.024	113	218768	47.410	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.820%	
50) Toluene-d8	10.446	98	905030	51.474	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.940%	
62) 4-Bromofluorobenzene	12.736	95	296726	50.696	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.400%	
Target Compounds						
						Qvalue
16) Acetone	4.301	43	11681	8.944	ug/l	# 80
20) Methylene Chloride	5.098	84	32397	7.143	ug/l	# 77
43) Isopropyl Acetate	8.568	43	37859	4.618	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081821\
Data File : VN068205.D
Acq On : 18 Aug 2021 20:08
Operator : JC/MD
Sample : M3449-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1

Quant Time: Aug 19 03:08:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

